# **Technical Specifications**

| Heads                          | Specifications   |
|--------------------------------|--|
| Dimension(mm)                  | 450 x 350 x 385 (LXWXH)  |
| Weight                         | 9.7 Kg   |
| Graphic display                | 7 inches with capacitive touch                                 |
| Patient name length            | 24 Characters  |
| Test result storage capacity   | 5000 test results  |
| Light source                   | Halogen lamp 12V /20W  |
| Photo detector                 | Silicon photodiode (range 300-1000nm)                          |
| Minimum Aspiration Volume      | 250 μL   |
| Interface                      | USB, ETHERNET, GSM, Bluetooth                                  |
| Spectral Range                 | 340-630nm  |
| Wavelength selection           | Automatic via 8-position filter wheel, 6 standard interference |
|                                | filter, 340nm, 405nm, 505nm, 546nm, 578nm, 630nm.              |
| Provision for optional Filters | Two  |
| Flow cell system               | 18 μL flow cell with 10 mm light path, inter changeable with   |
|                                | disposable macro, semi-micro or special optical glass cuvette  |
| Carry Over                     | Below 1% with 500 μL of Aspirated Sample.                      |
| Aspiration                     | With Peristatic Pump   |
| Printer                        | Graphic, Thermal printer                                       |
| Power                          | 115 to 260 VAC, 47 to 63 Hz                                    |
| Power Consumption              | 110 W  |
| Temperature control            | 25°C, 30°C, 37°C   |
| Temperature stabilizing time   | From 25°C to 37°C in 5 minutes.                                |
| Temperature Sensibility        | 0.1°C  |
| Measuring Range                | 0-3.5 OD   |
| Measuring Time                 | 3 Seconds  |
| Drift                          | Lower than 0.005 in OD per hour.                               |
| Photometric linearity          | R <sup>2</sup> >99%  |
| Battery                        | 14.8V / 5.7AH  |
|                                |  |

ISO 9001: 2015 EN ISO 13485: 2016

# **AGAPPE DIAGNOSTICS LTD.**

"Agappe Hills", Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.

TEL: + 91 484 2867000 | productcorp@agappe.in | www.agappe.com

YOUR BEST PARTNER IN DIAGNOSTICS

ADL/BR/PLUS/R1-03/19 Company reserves the right to change any desi

 $\epsilon$ 



Advanced Semi Automated Clinical Chemistry Analyser

# MISPA PLUS

**Advanced Semi Automated Clinical Chemistry Analyser** 


### **EXPERTISE THROUGH EXPERIENCE & INNOVATION**

Mispa PLUS is an elegant mix of Cutting-Edge technology, Innovation and Heritage. It is equipped with state-of-the-art Penta Lens Photometry system and IOT integration, perfectly designed to deliver high accuracy and ease of use. Delivering quality is our priority. Mispa Plus offers best performance to laboratories looking towards highest efficiency with optimal operative cost.

# PENTA LENS PHOTOMETRY SYSTEM ERA FLOW CELL EMF FILTER ETHERNET, BLUETOOTH & GSM CONNECTIVITY FOR ONLINE SYSTEM MANAGEMENT BUBBLE DETECTOR BATTERY BACKUP AUTO POWER SAVER MODE INBUILT INCUBATOR 7 INCH CAPACITIVE TOUCH SCREEN DISPLAY AUTO POWER SAVER MODE

### WHEN PERFORMANCE MATTERS

**MEMORY OF 5000 PATIENT RESULTS** 



**Internet Of Things** 

**Energy Matched Filter** 

**Bubble Detector** 

**In-Built Incubator** 

## MISPA PLUS

**Advanced Semi Automated Clinical Chemistry Analyser** 

### **ENERGY MATCHED FILTER (EMF)**

- Ensure long filter life
- Offers equal & stable intensity at all wavelengths for accurate estimation of test parameters.
- Offers high degree of optimization, precise and accurate results.

### **ERA FLOW CELL**

- Ensure clear light passage for accurate and reproducible results.
- Flow cell holder quickly achieves stable temperature for optimal reaction.
- Specially designed 18uL flow cell low sipping volume.
- Low aspirating volume of 250uL is only required for the test procedure.

### **BUBBLE DETECTOR**


- Eliminates error
- Ensures accurate and reliable testing result











**ENGINEERING EXCELLENCE - CASTING TRUST**